Citation for published version:

Lee, BJ, Sukri, NM, Ogden, H, Vine, C, Thake, CD, Turner, JE & Bilzon, JLJ 2015, 'A comparison of two commercially available ELISA methods for the quantification of human plasma heat shock protein 70 during rest and exercise stress', *Cell Stress Chaperones*, vol. 20, no. 6, pp. 917-926. <https://doi.org/10.1007/s12192-015-0610-3>

DOI:

[10.1007/s12192-015-0610-3](https://doi.org/10.1007/s12192-015-0610-3)

Publication date:

2015

Document Version

Peer reviewed version

[Link to publication](#)

Publisher Rights

CC BY

The final publication is available at Springer via <http://dx.doi.org/10.1007/s12192-015-0610-3>

University of Bath

Alternative formats

If you require this document in an alternative format, please contact:
openaccess@bath.ac.uk

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

A comparison of two commercially available ELISA methods for the quantification of human plasma heat shock protein 70 during rest and exercise stress

Background. This study compared resting and exercise heat/hypoxic-stress induced levels of plasma eHSP70 in humans using two commercially available ELISA kits.

Methods. EDTA plasma samples were collected from 21 males during two separate investigations. Participants in Part A completed a 60 min treadmill run in the heat (HOT70; 33.0 ± 0.1 °C, $28.7 \pm 0.8\%$, $n = 6$) at $70\% \text{VO}_{2\text{max}}$. Participants in Part B completed 60 minutes of cycling exercise at $50\% \text{VO}_{2\text{max}}$ in either hot (HOT50; 40.5°C , $25.4 \text{ RH}\%$, $n = 7$) or hypoxic (HYP50; $\text{F}_{\text{I}}\text{O}_2 = 0.14$, 21°C , $35\% \text{ RH}$, $n = 8$) conditions. Samples were collected prior to and immediately upon termination of exercise and analysed for eHSP70 using EKS-715 high sensitivity HSP70 ELISA, and new ENZ-KIT-101 AMP'D™ HSP70 high sensitivity ELISA. **Results.** ENZ-KIT was superior in detecting resting eHSP70 ($1.54 \pm 3.27 \text{ ng.mL}^{-1}$; range 0.08 to 14.01 ng.mL^{-1}), with concentrations obtained from 100% of samples compared to 19% with EKS-715 assay. The ENZ-KIT requires optimisation prior to running samples in order to ensure participants fall within the standard curve, a step not required with EKS-715. Using ENZ-KIT, a 1:4 dilution allowed for quantification of resting HSP70 in 26/32 samples, with a 1:8 ($n = 3$) and 1:16 ($n = 3$) dilution required to determine the remaining samples. After exercise eHSP70 was detected in 6/21 and 21/21 samples using EKS-715 and ENZ-KIT respectively. eHSP70 was increased from rest after HOT70 ($p < 0.05$), but not HOT50 ($p > 0.05$) or HYP50 ($p > 0.05$) when analysed using ENZ-KIT. **Conclusion.** It is recommended that future studies requiring the precise determination of resting plasma eHSP70 use the ENZ-KIT (i.e., HSP70 Amp'd® ELISA) instead of the EKS-715 assay, despite additional assay development time and cost required.

1 **Title:** A comparison of two commercially available ELISA methods for the
2 quantification of human plasma heat shock protein 70 during rest and exercise stress
3

4 **Running title:** Comparison of two ELISA methods for plasma HSP70 quantification.
5

6 **Key Words:** Heat shock proteins (HSP), acute heat stress, acute hypoxia, human
7

8 **Authors:** Lee BJ^{1,2}, Sukri, NM¹ Ogden H¹, Vine C¹, Thake CD², Turner JE¹, and
9 Bilzon JLJ¹.
10

11 1: Department for Health,
12 University of Bath,
13 Claverton Down,
14 Bath,
15 UK.
16

17 2: Department of Biomolecular and Sport Sciences,
18 Coventry University,
19 Priory Street,
20 Coventry,
21 UK.
22

23 **Correspondance.**

24 Ben Lee, Ph.D
25 The University of Bath
26 Department of Health
27 Claverton Down
28 Bath
29 BA2 7AY
30

31 **Telephone:** +44 (0)1225 383029

32 **Email:** B.J.Lee@bath.ac.uk
33
34

Abstract

Background. This study compared resting and exercise heat/hypoxic-stress induced levels of plasma eHSP70 in humans using two commercially available ELISA kits.

Methods. EDTA plasma samples were collected from 21 males during two separate investigations. Participants in Part A completed a 60 min treadmill run in the heat (HOT70; 33.0 ± 0.1 °C, $28.7 \pm 0.8\%$, $n = 6$) at 70% $\dot{V}O_{2\max}$. Participants in Part B completed 60 minutes of cycling exercise at 50% $\dot{V}O_{2\max}$ in either hot (HOT50; 40.5°C , 25.4 RH%, $n = 7$) or hypoxic (HYP50; $F_{I}O_2 = 0.14$, 21°C , 35% RH, $n = 8$) conditions. Samples were collected prior to and immediately upon termination of exercise and analysed for eHSP70 using EKS-715 high sensitivity HSP70 ELISA, and new ENZ-KIT-101 AMP'D™ HSP70 high sensitivity ELISA.

Results. ENZ-KIT was superior in detecting resting eHSP70 (1.54 ± 3.27 ng.mL⁻¹; range 0.08 to 14.01 ng.mL⁻¹), with concentrations obtained from 100% of samples compared to 19% with EKS-715 assay. The ENZ-KIT requires optimisation prior to running samples in order to ensure participants fall within the standard curve, a step not required with EKS-715. Using ENZ-KIT, a 1:4 dilution allowed for quantification of resting HSP70 in 26/32 samples, with a 1:8 ($n = 3$) and 1:16 ($n = 3$) dilution required to determine the remaining samples. After exercise eHSP70 was detected in 6/21 and 21/21 samples using EKS-715 and ENZ-KIT respectively. eHSP70 was increased from rest after HOT70 ($p < 0.05$), but not HOT50 ($p > 0.05$) or HYP50 ($p > 0.05$) when analysed using ENZ-KIT.

Conclusion. It is recommended that future studies requiring the precise determination of resting plasma eHSP70 use the ENZ-KIT (i.e., HSP70 Amp'd® ELISA) instead of the EKS-715 assay, despite additional assay development time and cost required.

Introduction

Heat shock proteins (HSPs) are an evolutionarily conserved family of proteins, with individual members named according to their molecular weight. Intracellular HSPs are expressed both constitutively and accumulate after exposure to a wide array of physiological and psychological stressors (Mosely 1996; Kregal 2002; Horowitz 2007). The 70 kilodalton (kDa) HSP (HSP70/HSPA1A, Kampinga et al. 2009) remains the most widely studied member of the HSP family due to its multiple functions related to de novo protein folding (Fink 1999), refolding (Hartl 1996), degradation (Garrido et al. 2001) and intracellular anti-inflammatory action observed

following induction (Ianaro et al. 2001) – all important factors in the maintenance of protein homeostasis.

In addition to its intracellular HSP70 (iHSP70) function, HSP70 has been detected in the circulation (i.e., plasma; Pockley et al. 1998), where it is described as an extracellular HSP (eHSP, Fleshner et al. 2003). Extracellular HSP70 stimulates neutrophil microbiocidal activity (Ortega et al. 2006), chemotaxis (Ortega et al. 2009), and induces cytokine production via a CD14 mediated pathway (Asea et al. 2000), thereby promoting innate immune activation (Krause et al. 2015). Elevated resting concentrations of eHSP70 has been positively correlated with insulin resistance (Krause et al. 2014), and disease progression in auto-immune (Luo et al. 2008) and inflammatory diseases (Schick et al. 2004; Najafizadeh et al. 2015). Therefore eHSP70 may be a useful biomarker when monitoring the progression of diseases in which low grade inflammation plays a role (Krause et al. 2012), such as sarcopenia (Ogawa et al. 2012), rheumatoid arthritis (Najafizadeh et al. 2015), diabetes (Krause et al. 2014), and obesity (Chung et al. 2008). The balance between the anti-inflammatory action of iHSP70, and pro-inflammatory action eHSP70, may determine the outcome (induction or attenuation of inflammation) during disease progression, in response to a treatment, or following an exercise bout (Krause et al. 2015). Therefore understanding the eHSP70 response and its actions is of importance in clinical situations.

In many studies the determination of basal eHSP70 has proven to be problematic (Ogawa et al. 2012; Lee et al. 2014; Najafizadeh et al. 2015) with no consensus normative data for resting eHSP70 reported to date. Sample handling considerations (e.g., blood collection tubes temporarily stored on ice *versus* room temperature, using Heparin *versus* EDTA as an anticoagulant) have been suggested as explanations underlying the variability in eHSP70 values reported in the literature. For example, serum appears to yield lower basal values than those obtained by plasma (Whitham and Fortes 2006), which may be an artifact of eHSP70 binding to proteins involved in the clotting process, such as fibrin and fibrinogen (Whitham and Fortes 2006). Additionally, EDTA is recommended for use over heparin as higher resting concentrations are derived (Whitham and Fortes, 2006). Despite the recommendation to use EDTA plasma for eHSP70 determination (Whitham and Fortes 2006), multiple

studies have attempted to quantify eHSP70 in serum for a range of conditions, with varied results (Dulin et al. 2010; Najafizadeh et al. 2015).

Another factor affecting the detection of eHSP70 in the circulation at rest could be attributed to the sensitivity and detection capabilities of the most widely cited enzyme linked immunosorbant assay (ELISA), 'EKS-715 HSP70 high sensitivity ELISA' (Enzo life sciences, Lausen, Switzerland). This assay, which has been cited by many papers investigating eHSP70 in humans, has a working range of 0.20 – 12.5 ng·mL⁻¹, a sensitivity of 0.09 ng·mL⁻¹, and is recommended by Cell Stress Society International (2011). In studies investigating eHSP70, resting values are highly variable, with many papers unable to detect basal eHSP70 in participants (Lee et al. 2014; Gibson et al. 2014), or concentrations reported at the lower portion of a standard curve (< 0.20 ng·mL⁻¹, Ogawa et al. 2012; Rodrigues-Krause et al. 2012; Gibson et al. 2014; Lee et al. 2014; Lee et al. 2015). This issue appears not to be uniform in the literature however, as some studies report concentrations that are much higher (17.0 ± 2.6 ng·mL⁻¹; Ruell et al. 2006).

A new ELISA has become available (ENZ-KIT-101-001 HSP70 Amp'd® ELISA), sensitive to 0.007 ng·mL⁻¹ with a working range of 0.039-5.00 ng·mL⁻¹. The increased sensitivity of this kit is mediated by an alkaline-phosphatase (AP) conjugate binding to a signal amplification substrate, which enhances colour production at lower analyte concentrations. The increased sensitivity therefore affords the potential to determine normative resting eHSP70 values for a range of individuals with a variety of conditions and could allow for a more sensitive determination of stress-induced changes in eHSP70.

Exercise stress can be used as a tool to study the eHSP70 response. Following exercise, both with and without a thermal component, eHSP70 is elevated in the circulation in a duration and intensity dependent manner (Whitham et al. 2007; Selkirk et al. 2009; Periard et al. 2013; Gibson et al. 2014; Lee et al. 2015). The magnitude of the post-exercise eHSP70 response has been related to a minimum endogenous requirement, which suggests that thresholds of core temperature, rate of core temperature change, and parasympathetic/sympathetic drive all play as of yet undetermined roles in the magnitude of this response (Gibson et al. 2014). A better

understanding of the criteria required to increase or decrease circulating eHSP70 may provide researchers with a useful biomarker for assessing therapeutic approaches to inflammation-related diseases, as well as improve understanding regarding eHSP70 function following acute exercise, and repeated periods of exercise, adaptation, and acclimation to extreme environments.

The aim of this investigation was to compare resting and exercise induced levels of EDTA plasma eHSP70 in Humans using the EKS-715 high sensitivity HSP70 ELISA and ENZ-KIT-101 Amp'd® HSP70 High Sensitivity ELISA methods. It was hypothesised that the use of the AP conjugate and signal amplification step would allow a more sensitive determination of basal eHSP70 within different cohorts of participants at rest while also showing greater sensitivity to different levels of “stress” induced by exercise undertaken in different environmental conditions.

Materials and Methods

Participants

Twenty-one recreationally active healthy males provided signed informed consent prior to participation in this study, which was granted approval by the NHS South-West Research Ethics Committee (Reference ID. 14/SW/0098, Part A) and Coventry University local ethics committee (Part B). All procedures were conducted in accordance with the principles outlined in the *Declaration of Helsinki*. Data reported in this investigation were collected from two larger experimental trials (Lee, 2014). The collected data represents a convenience sample of similarly characterised individuals providing EDTA treated plasma before and after a 60-minute bout of exercise under conditions of environmental stress.

All participants described themselves to be physically active, non-smokers with no prior history of cardiorespiratory illness. Participants were requested to abstain from caffeine (Lu et al. 2008) and alcohol consumption, as well as prolonged thermal exposures (baths, saunas, steam rooms, and tanning devices) for 72 hours prior to each laboratory visit, which were scheduled at similar times (08:30-09:30) between participants and trials. Participants adhered to an overnight fast prior to each trial and did not eat until after the final blood withdrawal.

Preliminary measurements

171 Participants in each part of the study were assessed for height, body mass and
172 percentage body fat in accordance with the International Society for the Advancement
173 of Kinanthropometry (ISAK) guidelines (Marfell-Jones et al. 2006).

174 Participants in Part A (n = 6) of the investigation completed a continuous incremental
175 running test to volitional exhaustion on a motorised treadmill (Woodway ELG70,
176 Weiss, Germany). The test protocol was modified from that of Taylor et al. (1955)
177 and was performed in thermoneutral conditions (19.7 ± 0.7 °C, $46.3 \pm 4.0\%$ RH).
178 After a five-minute warm-up at $6 \text{ km}\cdot\text{h}^{-1}$, the test began at a speed of $10 \text{ km}\cdot\text{h}^{-1}$ on a
179 1% inclination. Speed was then increased by $1 \text{ km}\cdot\text{h}^{-1}$ every three minutes until
180 reaching $13 \text{ km}\cdot\text{h}^{-1}$, when inclination was increased by 2% every two minutes.
181 Participants were instructed to run for as long as possible and signal when they felt
182 they could only complete one more minute to allow for a final set of recordings. Peak
183 oxygen consumption was determined for participants in Part B using an incremental
184 exercise test to volitional exhaustion on calibrated SRM cycle ergometer (n = 15,
185 Table 1) Schoberer Rad Meßtechnik, Welldorf, Germany). Resting blood lactate
186 (Biosen C-Line analyser, EKF Diagnostics, Germany) was determined from a finger
187 capillary whole blood sample following a 10-minute seated rest period. The test began
188 at a workload of 70W for 4-minutes and was then increased by 35W every 4 minutes
189 until a blood lactate value of $> 4 \text{ mmol}\cdot\text{L}^{-1}$ was reached. Thereafter, workload
190 increased 35W every 2 minutes until volitional exhaustion. A cadence of $70 \text{ rev}\cdot\text{min}^{-1}$
191 was maintained throughout. In both Part A and Part B, expired gases were collected
192 using 200L Douglas bags (Cranlea & Co, Birmingham, UK) during the final minute
193 of each stage. Heart rate (Polar FT1, Polar Electro OY, Kempele, Finland) and
194 perceived exertion (Borg 1976) were measured at the end of each gas collection.
195 Respiratory gas analysis was completed as previously described (Lee et al., 2014, Lee
196 et al., 2015). Peak oxygen consumption was considered to be achieved if two of the
197 following criteria were met: i) a respiratory exchange ratio of >1.1 , ii) a heart rate
198 greater than 95% of age predicted maximum ($220 - \text{age}$) and iii) a final blood lactate
199 value in excess of $8 \text{ mmol}\cdot\text{mL}^{-1}$.

200 **Experimental design**

Samples for eHSP70 analysis were obtained from two separate experiments which both involved a resting measure of eHSP70 and a measurement collected immediately after a 60-minute bout of exercise.

Part A

Samples were obtained before and immediately after a 60 minute treadmill run at a speed equivalent of 70% $\dot{V}O_{2max}$ (HOT70) from 6 healthy males (mean $\pm$ SD; age 20 $\pm$ 2 years; height 1.79 $\pm$ 0.04 meters; body mass 71.8 $\pm$ 2.7 kg; % body fat, 11.8 $\pm$ 3.3%; $\dot{V}O_{2max}$ 57.9 $\pm$ 9.7 mL.kg⁻¹.min⁻¹). All trials were performed in an environmental chamber that was regulated at a dry bulb temperature of 33.0 $\pm$ 0.1°C and relative humidity (RH) of 28.7 $\pm$ 0.8%, with blood samples obtained at rest, upon termination of exercise. Participants (n = 6) returned to the lab on two more occasions each separated by 14 days, to provide 2 further resting samples, which formed part of a larger experiment. These resting samples were included in the present analysis of resting data.

Part B

Samples were obtained before and immediately after a 60 minutes of cycling at a power output equivalent of 50% $\dot{V}O_{2max}$ in either hot (HOT50; 40.5°C, 25.4 RH%) or hypoxic (HYP50; F_IO₂ of ~0.14, 21°C, 35% RH) conditions. The hypoxic environment was generated by an oxygen filtration device (Hypoxico HYP-123 hypoxicator, New York, NY, USA) set to produce the desired F_IO₂. Participant characteristics for HOT50 (n = 7) were (mean $\pm$ SD): age = 22 $\pm$ 5 years; height 1.76 $\pm$ 0.05 meters; body mass 70.9 $\pm$ 5.7 kg; % body fat 13.2 $\pm$ 4.0%, $\dot{V}O_{2max}$ 54.9 $\pm$ 3.2 mL.kg⁻¹.min⁻¹. Participant characteristics for HYP50 were: 23.4 $\pm$ 4 years; height 1.80 $\pm$ 0.08 meters; body mass 70.0 $\pm$ 9.1 kg; % body fat 12.6 $\pm$ 3.7 %; $\dot{V}O_{2max}$ 52.2 $\pm$ 3.3 mL.kg⁻¹.min⁻¹.

Participant Preparation

Participant preparation and physiological measurements were completed in the same manner and at the same time intervals for both Part A and B. Prior to each visit, participants adhered to an overnight fast (Febbraio et al. 2002) and consumed 500 ml of plain water one hour before in accordance with the American College of Sports Medicine position stance on hydration (Sawka et al. 2007). Upon arrival, participants

began by voiding their bladder to provide a sample for hydration assessment via urine specific gravity (USG; Atago Refractometer, Jencons Pls, Leighton Buzzard, UK) and urine osmolarity (U_{osmo} ; Advanced 3300 Micro-Osmometer, Advanced Inc, Massachusetts, USA). Euhydration was assumed for urine specific gravity values of $\leq 1.020 \text{ g}\cdot\text{ml}^{-1}$ and osmolarity values of $\leq 700 \text{ mOsm}\cdot\text{kg}^{-1}$ (Armstrong et al. 1994). This control was not violated by any participant during any trial. Following this, participants measured their own nude body mass (Seca 880, Seca, Hamburg, Germany), inserted a calibrated rectal thermistor probe (Grant Squirrel 2020, Grant Instruments, Shepreth, UK) to a depth of 10 cm, and fitted a telemetric heart rate monitor around their chest (Polar FT1, Polar Electro OY, Kempele, Finland). An indwelling cannula (BD Insyte-W, Becton Dickinson, Utah, USA) was then inserted 2.5 cm into an antecubital vein of the participants left arm. After a 20 minute stabilisation period with the participant lying supine, a baseline 10 ml blood sample was then drawn, with patency of the cannula being maintained with saline (0.9% sodium chloride, Braun, Melsungen, Germany).

Physiological Measurements

Participants entered the regulated environmental chamber at 09:30. The exercise bout began with a standardised five minute warm-up, running on a motorised treadmill at a speed calculated to elicit a work rate of 50% $\dot{V}O_{2max}$ on a fixed 1% inclination (Jones & Drust 1996) in Part A, or a 15 minute seated wash-in for the hypoxic gas in group B. Upon completion of the warm-up/wash-in period, participants in Part A began a 60 minute run at a work rate of 70% $\dot{V}O_{2max}$ and participants in Part B began 60 minutes of cycling exercise at 50% $\dot{V}O_{2max}$ in the prescribed environmental conditions (HOT50 or HYP50).

During exercise, heart rate (HR), rectal temperature (T_{rectal}), ratings of perceived exertion (RPE) and thermal sensation (ISO, 1995, Part A, Part B) were all recorded at ten-minute intervals. The Physiological Strain Index (PSI) was subsequently calculated at each time point using heart rate and rectal temperature data as described by Moran et al. (1998). The T_{rectal} area under curve was calculated using a modification of the trapezium rule (Hubbard et al. 1977) when T_{rectal} exceeded 38.5°C (Cheuvront et al. 2008) and 39.0°C. A T_{rectal} of 38.5°C was selected as a possible threshold for eHSP70 appearance (Gibson et al. 2014). In instances where participants

266 did not complete the full 60 minute run/cycle, termination time was recorded and all
267 aforementioned measures taken in the final minute before cessation.

268 **Determination of extracellular HSP70**

269 Circulating eHSP70 was assessed using two commercially available ELISAs, EKS-
270 715 high sensitivity HSP70 kit
271 (http://static.enzolifesciences.com/fileadmin/files/manual/ADI-EKS-715_insert.pdf;
272 hereafter referred to as EKS-715) and ENZ-KIT-101-001 Amp'd® HSP70 high
273 sensitivity ELISA kit ([http://static.enzolifesciences.com/fileadmin/files/manual/ENZ-](http://static.enzolifesciences.com/fileadmin/files/manual/ENZ-KIT-101_insert.pdf)
274 [KIT-101_insert.pdf](http://static.enzolifesciences.com/fileadmin/files/manual/ENZ-KIT-101_insert.pdf), hereafter referred to as ENZ-KIT) according to the
275 manufacturer's instructions (Enzo Lifesciences, Lausen, Switzerland).

276
277 The ENZ-KIT is designed to replace traditional alkaline phosphatase substrates, such
278 as pNPP (p-Nitrophenyl phosphate), with a combination substrate and amplifier
279 system that results in greater sensitivity when compared to a classic substrate ELISA.
280 In the ENZ-KIT, bound AP converts a substrate that is utilized in a second enzyme
281 reaction system which is initiated by addition of the amplifier reagent. Figure 1
282 shows typical standard curves (n = 4) prepared on the same 96 well plate (HSP70
283 Clear Miotiter plate, catalogue number: 80-1581), using the same HSP70 high
284 sensitivity standard (Cat no: 80-1776). HRP conjugate (Cat no: 80-1778) was added
285 to HSP70 high sensitivity antibody (Cat no: 80-1777) and a TMB substrate (Cat no:
286 80-0350) used to develop the EKS-715. For ENZ-KIT, an AP conjugate (Cat no: 80-
287 2600) was added to the HSP70 high sensitivity antibody (Cat no: 80-1777) and
288 incubated with signal amplification substrate (Cat no: 80-2596) containing NADPH
289 prior to a final amplification step (Cat no: 80-2598). The amplification step allows for
290 greater (amplified) colour production at lower analyte concentrations resulting in an
291 increased assay sensitivity (Figure 1).

292
293 The EKS-715 kit has a sensitivity of 0.090 ng mL^{-1} and a working range of 0.20 to
294 12.5 ng mL^{-1} . The ENZ-KIT assay is sensitive to 0.007 ng mL^{-1} with a working range
295 of $0.039\text{--}5.00 \text{ ng mL}^{-1}$. Following an initial analysis of samples (one ENZ-KIT assay),
296 it became apparent that the minimum recommended dilution of 1:4 was not sufficient
297 in all cases, with some samples containing more HSP70 than the top standard. Thus a
298 further analysis using 1:4, 1:8 and 1:16 dilution step with assay diluent (sodium

carbonate) was necessary to determine the optimal dilution for each sample, with results multiplied by this dilution factor in order to give eHSP70 values in ng mL⁻¹. Once the optimal sample dilution was determined for each participant on the ENZ-KIT, pre and post exercise samples were analysed in duplicate.

Statistical analysis

A total of 32 resting blood samples from 21 individuals (11 repeat samples) were analysed for basal eHSP70. The between-assay co-efficient of variation was determined using standard concentration curves of 4 separate kits run using both EKS-715 and ENZ-KIT. The reliability of the ENZ-KIT assay was further assessed by comparing resting data obtained during a serial dilution test to the data obtained from a further assay on these samples measuring both pre and post exercise data. Where each ELISA method provided eHSP70 for paired samples, or samples assayed on separate occasions, Pearson correlations determined the relationship between each measurement. Mean and peak physiological, thermoregulatory and eHSP70 responses were analysed between groups using a one-way analysis of variance (ANOVA), and Tukey's honestly different test to explore main effects. All data analysis was performed using PASW software version 20.0 for Mac (SPSS, Chicago, IL, USA). Stepwise multiple regression analysis was performed using the three dependent variables (time spent above T_{rectal} 38.5°C, rate of change in T_{rectal}, and AUC for T_{rectal} 38.5°C) that were significantly correlated to post exercise eHSP70 concentrations. The significance level was set at $p < 0.05$ for all analysis. Data are reported as means $\pm$ SD unless otherwise stated and individual data shown where possible.

Results

pNPP Conjugate (EKS-715) compared with AP conjugate and amplifier substrate (ENZ-KIT)

Figure 1 illustrates a typical standard curve when the pNPP conjugate and TMB substrate is used (closed circles) in comparison to the increased sensitivity obtained from the AP conjugate and amplifier solution (open circles). The intra-assay precision, obtained by determining the coefficient of variation between duplicate samples obtained from the standard curves on 4 separate plates, was 2.6% and 4.1%

for EKS-715 and ENZ-KIT respectively. These data are lower than the manufacturer reported intra-assay precision of between 3.9 and 11.4% for EKS-715, and 7 and 15% for the ENZ-KIT. Inter-assay precision was also determined from the standard curves of 4 separate assays performed on 4 separate occasions, and was 4.9% and 6.2% for EKS-715 and ENZ-KIT respectively. These values were also lower than manufacturer reported inter-assay variation (EKS-715 = 12.8 – 19.1%; ENZ-KIT = 7.7 – 9.7%).

Extracellular HSP70 at rest

The ENZ-KIT was able to detect basal eHSP70 in all 32 resting samples (Figure 2, $1.54 \pm 3.27 \text{ ng}\cdot\text{mL}^{-1}$). In contrast, the EKS-715 assay did not detect eHSP70 in 26 out of 32 resting samples analyzed (81%). When results were available from both kits ($n = 6$), there was a good correlation between values ($r = 0.86$, $p = 0.0004$, Figure 2), with values not significantly different between kits ($t = 0.35$, $p = 0.72$). In 15 of the samples measured with ENZ-KIT (47%), eHSP70 was below the $0.20 \text{ ng}\cdot\text{mL}^{-1}$ limit of EKS-715 standard curve ($0.15 \pm 0.04 \text{ ng}\cdot\text{mL}^{-1}$; 95% CI = 0.13 to $0.17 \text{ ng}\cdot\text{mL}^{-1}$).

A minimum dilution of 1:4 (sample to assay diluent) is recommended to remove matrix interference during the ENZ-KIT assay. In the present investigation we found the 1:4 dilution allowed for determination of basal HSP70 in 26/32 samples studied. For samples with resting concentrations of eHSP70 above the top standard concentration ($5.00 \text{ ng}\cdot\text{mL}^{-1}$), a 1:8 dilution ($n = 3$) and 1:16 ($n = 3$) were required to locate data on the standard curve. No participants exhibited eHSP70 values below the detection limit of $0.039 \text{ ng}\cdot\text{mL}^{-1}$ using the ENZ-KIT

A further determination of ENZ-KIT assay reliability was made by comparing resting eHSP70 data obtained from the serial dilution plate, to the resting data obtained during the test run on a separate plate (Part A: $n = 6$, Part B $n = 15$; $r = 0.998$, $p < 0.001$, Figure 2 Panel E and F), indicating good inter-assay reliability (CV = 7.9%).

Physiological and thermoregulatory responses to each stressor

The duration of exercise undertaken at 70% $\dot{V}\text{O}_2\text{max}$ (i.e., HOT70, 54.0 ± 9.4 minutes) was shorter than the duration of exercise undertaken at 50% $\dot{V}\text{O}_2\text{max}$ (i.e.,

365 HOT50, HYP50, 60 ± 0.0 minutes; $f_{(2, 18)} = 4.25$, $p = 0.032$). Physiological and
366 thermoregulatory data are shown in Table 1.

367

368 Although some thermoregulatory responses (peak, delta and rate of T_{rectal} change)
369 were greater in HOT50 compared to HYP50 (Table 1), no other differences in
370 physiological response (e.g. mean and peak HR and PSI) were observed between
371 these conditions. In contrast, greater mean and peak exercising HR and T_{rectal}
372 responses were observed in HOT70 compared to HOT50 and HYP50. Similarly, the
373 delta change in T_{rectal} , the rate of T_{rectal} change, the duration of the exercise bout spent
374 above both 38.5°C and 39.0°C and AUC for these temperatures were all greater in
375 HOT70 compared to HOT50 and HYP50 ($p < 0.05$; Table 1). The data therefore
376 indicate two different levels of physiological strain were achieved (HOT70 versus
377 HOT50 and HYP50).

378

379 **The exercise-induced eHSP70 response**

380 In accordance with the resting data, EKS-715 only detected post exercise eHSP70 in
381 the 6 samples that had detectable eHSP70 at rest, with 5 samples obtained in HYP50
382 and 1 sample from the HOT50. Post exercise eHSP70 obtained from the EKS-715 kit
383 ($n = 6$, 3.92 ± 4.34 ng.mL⁻¹) had a good relationship to those obtained with the ENZ-
384 KIT ($n = 6$, 3.37 ± 5.38 ng.mL⁻¹, $r = 0.84$).

385

386 There was a significant group $\times$ time interaction ($F_{(2, 17)} = 4.235$, $p = 0.03$) for
387 eHSP70 when analyzed using ENZ-KIT. Resting HSP70 was higher in HOT70 than
388 HOT50 or HYP50 group ($p < 0.05$; Figure 3).

389

390 Exercise results in an increase in eHSP70 from 2.79 ± 2.59 ng.mL⁻¹ (95% CI = 0.074
391 to 5.51 ng.mL⁻¹) at rest to 3.51 ± 2.90 ng.mL⁻¹ (0.47 to 6.56 ng.mL⁻¹) in the HOT70
392 group ($t = 3.82$, $p = 0.012$). However, with the HOT50 trial, eHSP70 was unchanged
393 from 0.22 ± 0.13 ng.mL⁻¹ (95% CI = 0.084 to 0.362 ng.mL⁻¹) at rest to 0.20 ± 0.16
394 ng.mL⁻¹ (95% CI = 0.034 to 0.370 ng.mL⁻¹) following exercise ($t = 0.886$ $p = 0.410$).
395 In addition, in the HYP50 trial, eHSP70 was unchanged from 2.82 ± 5.51 ng.mL⁻¹
396 (95% CI = 2.96 to 8.60 ng.mL⁻¹) at rest to 2.85 ± 5.56 ng.mL⁻¹ (95% CI = 2.98 to 8.69
397 ng.mL⁻¹) following exercise ($t = 0.635$, $p = 0.545$). Thus only HOT70 induced
398 changes of eHSP70 above resting values (Figure 3).

Relationship between eHSP70 and thermo-physiological measures

Time spent above 38.5°C ($r = 0.54$), rate of change in T_{rectal} ($r = 0.52$), and AUC for T_{rectal} 38.5°C ($r = 0.47$) were entered into a stepwise multiple regression analysis to assess the association of these variables in post exercise eHSP70 concentration. The only predictor variable was the duration of exercise above 38.5°C. The adjusted R^2 for this model was 0.26 with a large standard error of 55.1.

Discussion

The aim of this study was to compare two commercially available high sensitivity ELISAs for the determination of eHSP70 in plasma. The results illustrate that the ENZ-KIT (Enzo Lifesciences, Lausen, Switzerland) is more sensitive than the EKS-715 (Enzo Lifesciences, Lausen, Switzerland) when quantifying both resting and post exercise eHSP70 values in a sample of healthy, moderately trained males. The increased sensitivity and lower working range, facilitated by the use of amplifier reagents, significantly improves the ability of the ENZ-KIT assay to detect resting eHSP70 in plasma thereby supporting our hypothesis.

In the present investigation only 6 of the 21 samples analysed using the EKS-715 kit allowed for the quantification of eHSP70 at rest and after exercise, whereas ENZ-KIT provided data for all resting and all post exercise samples (Figure 2, Panel B). The increased sensitivity of ENZ-KIT introduces the potential requirement for serial dilution of samples to ensure results are not above the standard curve, thereby reducing the reliance on extrapolation. In the current analysis, the manufacturers recommended minimum sample dilution of 1:4 was sufficient to detect eHSP70 in 26/32 samples. Additional dilutions of 1:8 ($n = 3$) and 1:16 ($n = 3$) were necessary in the instances when data fell above the standard curve. It may be prudent for researchers using the ENZ-KIT to conduct a 1:4 and 1:8 serial dilution of all resting samples prior to full analysis to ensure that all samples can be analysed together, potentially saving both time and the additional cost of running more assays. The ENZ-KIT demonstrated excellent reproducibility between individual assay kits, with duplicate measurements of resting values highly correlated between assays ($R^2 = 0.99$, Figure 2, Panel F).

Heat shock proteins play an important role in maintaining cellular protein homeostasis, with HSP dysfunction implicated in the pathology of Alzheimer's disease, Parkinson's disease, cardiovascular disease, and sarcopenia (Krause et al. 2015). It is therefore surprising that there is currently no substantial normative data regarding resting eHSP70 for either healthy individuals or clinical cohorts. Indeed, a characteristic of eHSP70 research is the large variability both between studies (Ruell et al. 2006; Whitham and Fortes 2006; Gibson et al. 2014; Lee et al. 2015) and within studies, likely exacerbated by small samples sizes typically used in exercise studies (Gibson et al. 2014; Lee et al. 2015). Between-participant variation is evident in the present investigation, in which three groups of seemingly physiologically-matched males present significantly different eHSP70 values at rest (Figure 3). Such a large disparity in resting data will limit the ability to detect changes in eHSP70 concentrations between two or more matched groups. When a repeated measures design is not feasible it may be appropriate for experimenters to match individuals based on resting eHSP70 concentrations rather than more common physiological and anthropometric features (providing that eHSP70 is an important outcome). Doing so may facilitate a clearer understanding regarding responders, such as the participant with high post HOT50 eHSP70 concentration (Figure 3, Panel B; rest = 0.67 ng.mL⁻¹, post exercise = 2.29 ng.mL⁻¹), and non-responders to thermal or hypoxic stress.

The reasons for the observed disparity between similarly matched groups is unclear and could not be determined in the present work. In order to elucidate the role this biochemical marker plays in health and disease future studies should aim to thoroughly characterize an individual's lifestyle factors, normal weekly physical activity levels, and anthropometric and physiological characteristics. The increased sensitivity of the ENZ-KIT compared to other commercially available kits allow for a sensitive quantification of resting eHSP70 across a wide range of populations. Only once these data have been collected in a sufficiently large sample can eHSP70 be investigated as a potential biomarker of, for example, sarcopenia (Ogawa et al. 2012). Some studies have attempted to determine the efficacy of eHSP70 in clinical scenarios using the EKS-715 assay. Based on the present work, it is likely that prior results have been influenced by sensitivity. For example Ogawa et al. (2012) conducted a detailed study in which 652 elderly Japanese males and females were screened for a range of biochemical and physiological markers related to sarcopenia,

such as TNF- α , interleukin-6 (IL-6) and C-reactive protein (CRP). However the majority of the eHSP70 data reported (436/652) was near to ($n = 207$; $0.13 - 0.22$ ng.mL⁻¹), or below ($n = 229 < 0.13$ ng.mL⁻¹) the EKS-715 assay standard curve (0.20 ng.mL⁻¹). Thus, although these results are important, the present work highlights that the data should be interpreted with caution. The data in the present investigation has demonstrated that, even where eHSP70 values are theoretically within the measurable range of the EKS-715 assay (e.g. concentrations of over 0.20 ng.mL⁻¹), they are not always detectable by EKS-715, but were all detected with ENZ-KIT (Figure 2). Thus, when resting eHSP70 concentrations are required for clinical observations or for studying the role of eHSP70 in health and disease, the use of the ENZ-KIT is recommended above EKS-715.

Exercise is known to increase concentrations of eHSP70 in an intensity and duration dependent manner, with a 60 minute run at $75\% \dot{V}O_{2\max}$ leading to a 175% increase in eHSP70, compared to a 140% increase when running for 120 minutes at $60\% \dot{V}O_{2\max}$ (Fehrenbach et al. 2005). The sum exercise stress of cycling is less than of running due to its low impact and less muscle-damaging nature, thus cycling induced elevations in eHSP70 concentration are much lower than those observed following running (Febbraio et al. 2002; Febbraio et al. 2004; Lancaster et al. 2004). Post exercise increases in eHSP70 have been hypothesized to relate to a minimum endogenous level of thermal strain, corresponding to a T_{rectal} of $> 38.5^{\circ}\text{C}$ (Amorim et al. 2008; Gibson et al. 2014). Both the rate of T_{rectal} increase, and change in T_{rectal} are also thought to be important factors affecting eHSP70 concentrations after a stressor (Periard et al. 2012; Gibson et al. 2014).

In addition to the ability of the ENZ-KIT to quantify resting eHSP70, we therefore examined the post-exercise response at 3 levels of physiological strain using two distinctly different modes (running and cycling) of hyperthermic exercise, as well as hypoxic stress (Table 1). The data presented in Table 1 indicate two levels of physiological strain were achieved, with HOT70 eliciting significantly greater mean and peak HR, T_{rectal} and PSI compared to HOT50 and HYP50. No significant differences between the HOT50 and HYP50 groups were observed, agreeing with previous work showing exercise at $50\% \dot{V}O_{2\max}$ in either 40°C heat or an $F_{\text{I}O_2}$ of 0.14

to be of a comparable physiological stress up to 40 minutes of exercise (Lee et al. 2014). The 60 minute duration of exercise used herein was therefore insufficient for significant differences in physiological strain between HOT50 and HYP50 to become apparent (Girard and Racinais, 2014; Lee et al. 2014).

As with the resting data, the EKS-715 kit only detected eHSP70 in the 6 samples that also provided resting data ($n = 5$ from HYP50, and $n = 1$ from HOT50). In contrast a 52% increase (Pre, 2.79 ± 2.58 ; Post, 3.51 ± 2.90 ng.mL⁻¹) in eHSP70 was observed immediately after HOT70 when samples were analyzed with ENZ-KIT. This is comparable with other studies using a similar level of external heat stress at a lower exercise intensities (e.g. Periard et al. 2012) but for a longer duration (90 minutes, Gibson et al. 2014), and further illustrates the utility of the ENZ-KIT over EKS-715.

The higher post exercise eHSP70 concentrations reported in the present investigation were observed despite the shorter duration of exposure (54.0 ± 9.3 minutes compared to 90.0 ± 0.0 minutes in Gibson et al. 2014) and are likely due to the increased exercise intensity (70% vs 50% $\dot{V}O_{2max}$ in Gibson et al. 2014), and different exercise mode (running Vs. cycling) used between investigations. Additionally, thermoregulatory stress, evidenced by the AUC for a T_{rectal} of $>38.5^{\circ}C$ ($9.21 \pm 1.95^{\circ}C.min^{-1}$), the duration spent above $38.5^{\circ}C$ (17.5 ± 10.4 minutes) and duration spent above $39^{\circ}C$ (8.9 ± 1.6 minutes), mean and peak T_{rectal} , change in T_{rectal} and rate of change in HOT70 were all higher than those reported after cycling exercise at 50% $\dot{V}O_{2max}$ for 90 minutes at $30.2^{\circ}C$, 51% RH (Gibson et al. 2014). Thus the total level of endogenous strain was greater in the present investigation, and is reflected in the eHSP70 results.

In contrast, no post exercise increase was observed following HOT50 or HYP50 (Figure 3). It is likely that while the thermal component of HOT50 was sufficient for increased eHSP70, the duration of exercise, and therefore overall level of exogenous strain, was not sufficient to increase eHSP70. Of the suggested endogenous requirement for post exercise increases in eHSP70 (peak T_{rectal} of $>39.2^{\circ}C$, a mean T_{rectal} of $38.6^{\circ}C$ for a period of ~57 minutes, a core temperature change of $2.2^{\circ}C$ from baseline at a rate of $1.6^{\circ}C h^{-1}$ and a mean heart rate of 153 bt.min⁻¹), the HOT50 trial only achieved the required heart rate, and the HYP50 group failed to reach any of

these potential eHSP70 inducing thresholds. Our data therefore lend support to the notion of a minimum endogenous criteria required for eHSP70 induction (Amorim et al. 2008; Gibson et al. 2014). The increased sensitivity afforded by the ENZ-KIT may allow for a more detailed and nuanced description of minimum eHSP70 inducing criteria during and after exercise stress in future studies.

In summary, this investigation presents preliminary data showing the effectiveness of the ENZ-KIT assay for detecting and quantifying resting eHSP70 at the low end of the measurable range in young healthy males. Secondly, our results support the notion that a minimum endogenous strain threshold needs surpassing in order to increase systemic HSP70. As a result, it is recommended that all future investigations requiring accurate resting eHSP70 quantification use the ENZ-KIT assay in place of EKS-715. The increased sensitivity afforded by this assay could provide a more in depth understanding of normal and abnormal levels of systemic eHSP70, and provide novel information regarding the use of eHSP70 as a biomarker of disease.

Abbreviations

AUC; Area under the curve. **AP**; Alkaline phosphatase. **ENZ-KIT-101**; Amp'd® HSP70 High Sensitivity ELISA kit. **eHSP70**; Extracellular heat shock protein 70. **EKS-715**; HSP70 high sensitivity ELISA kit. **ELISA**; Enzyme linked immunosorbant assay. **FiO₂**; fraction of inspired oxygen. **HEAT70**; Heat running trial. **HEAT50**; heat cycling trial **HSP70**; Heat shock protein 72. **HR**; Heart rate. **HYP**; Hypoxic trial. **IL-10**; Interleukin 10. **kDa**; KiloDalton. **pNPP**; p-Nitrophenyl phosphate. **PSI**; Physiological strain index. **RPE**; rating of perceived exertion. **T_{rectal}**; Rectal temperature. **RH**; Relative humidity. **SpO₂**; Arterial oxygen saturation. **TS**; Thermal sensation. **TNF- α** ; Tumour necrosis factor alpha. **$\dot{V}O_{2max}$** ; Maximal oxygen consumption.

Acknowledgements

The authors would like to thank Morgan Mathieu, Ph.D, for his technical advice during the optimization steps of the ELISA assays. We would also like to thank all the volunteers for their time during the study.

Conflict of Interest

The authors declare no conflict of interests

References

- AMORIM, F. T., YAMADA, P. M., ROBERGS, R. A., SCHNEIDER, S. M. & MOSELEY, P. L. 2008. The effect of the rate of heat storage on serum heat shock protein 72 in humans. *European journal of applied physiology*, 104, 965-972.
- ARSLAN, M. A., CSERMELY, P. & SÖTI, C. 2006. Protein homeostasis and molecular chaperones in aging. *Biogerontology*, 7, 383-389.
- ASEA, A. 2005. Stress proteins and initiation of immune response: chaperokine activity of hsp72. *Exercise immunology review*, 11, 34.
- ASEA, A. 2006. Initiation of the immune response by extracellular Hsp72: chaperokine activity of Hsp72. *Current immunology reviews*, 2, 209.
- BORG, G. 1976. Simple rating methods for estimation of perceived exertion. *Physical work and effort*, 39-46.
- CHEUVRONT, S. N., CHINEVERE, T. D., ELY, B. R., KENEFICK, R. W., GOODMAN, D. A., MCCLUNG, J. P. & SAWKA, M. N. 2008. Serum S-100 beta response to exercise-heat strain before and after acclimation. DTIC Document.
- DRUST, B., WATERHOUSE, J., ATKINSON, G., EDWARDS, B. & REILLY, T. 2005. Circadian rhythms in sports performance-an update. *Chronobiology international*, 22, 21-44.
- DULIN, E., GARCÍA-BARRENO, P. & GUIASOLA, M. C. 2010. Extracellular heat shock protein 70 (HSPA1A) and classical vascular risk factors in a general population. *Cell Stress and Chaperones*, 15, 929-937.
- FEBBRAIO, M. A., MESA, J. L., CHUNG, J., STEENSBERG, A., KELLER, C., NIELSEN, H. B., KRUSTRUP, P., OTT, P., SECHER, N. H. & PEDERSEN, B. K. 2004. Glucose ingestion attenuates the exercise-induced increase in circulating heat

601 shock protein 72 and heat shock protein 60 in humans. *Cell stress & chaperones*, 9,
602 390.
603
604 FEBBRAIO, M. A., STEENBERG, A., WALSH, R., KOUKOULAS, I., HALL, G.
605 V., SALTIN, B. & PEDERSEN, B. K. 2002. Reduced glycogen availability is
606 associated with an elevation in HSP72 in contracting human skeletal muscle. *The*
607 *Journal of Physiology*, 538, 911-917.
608
609 FEHRENBACH, E., NIESS, A., VOELKER, K., NORTHOFF, H. & MOOREN, F.
610 2005. Exercise intensity and duration affect blood soluble HSP72. *International*
611 *journal of sports medicine*, 26, 552-557.
612
613 FINK, A. L. 1999. Chaperone-mediated protein folding. *Physiological reviews*, 79,
614 425-449.
615
616 GARRIDO, C., GURBUXANI, S., RAVAGNAN, L. & KROEMER, G. 2001. Heat
617 shock proteins: endogenous modulators of apoptotic cell death. *Biochemical and*
618 *biophysical research communications*, 286, 433-442.
619
620 GIBSON, O. R., DENNIS, A., PARFITT, T., TAYLOR, L., WATT, P. W. &
621 MAXWELL, N. S. 2014. Extracellular Hsp72 concentration relates to a minimum
622 endogenous criteria during acute exercise-heat exposure. *Cell Stress and Chaperones*,
623 19, 389-400.
624
625 GIRARD, O. & RACINAIS, S. 2014. Combining heat stress and moderate hypoxia
626 reduces cycling time to exhaustion without modifying neuromuscular fatigue
627 characteristics. *European journal of applied physiology*, 114, 1521-1532.
628
629 GOLDBERG, A. L. 2003. Protein degradation and protection against misfolded or
630 damaged proteins. *Nature*, 426, 895-899.
631
632 HARTL, F. U. 1996. Molecular chaperones in cellular protein folding. *Nature*, 381,
633 571-9.
634

635 HICKMAN-MILLER, H. D. & HILDEBRAND, W. H. 2004. The immune response
636 under stress: the role of HSP-derived peptides. *Trends in immunology*, 25, 427-433.
637

638 HOROWITZ, M. 2007. Heat acclimation and cross-tolerance against novel stressors:
639 genomic–physiological linkage. *Progress in brain research*, 162, 373-392.
640

641 HUBBARD, R., BOWERS, W., MATTHEW, W., CURTIS, F., CRISS, R.,
642 SHELDON, G. & RATTEREE, J. 1977. Rat model of acute heatstroke mortality.
643 *Journal of Applied Physiology*, 42, 809-816.
644

645 IANARO, A., IALENTI, A., MAFFIA, P., PISANO, B. & DI ROSA, M. 2001.
646 HSF1/hsp72 pathway as an endogenous anti-inflammatory system. *FEBS letters*, 499,
647 239-244.
648

649 JOHNSON, J. D. & FLESHNER, M. 2006. Releasing signals, secretory pathways,
650 and immune function of endogenous extracellular heat shock protein 72. *Journal of*
651 *Leukocyte Biology*, 79, 425-434.
652

653 JONES, A. M. & DOUST, J. H. 1996. A 1% treadmill grade most accurately reflects
654 the energetic cost of outdoor running. *Journal of sports sciences*, 14, 321-327.
655

656 KAMPINGA, H. H., HAGEMAN, J., VOS, M. J., KUBOTA, H., TANGUAY, R. M.,
657 BRUFORD, E. A., CHEETHAM, M. E., CHEN, B. & HIGHTOWER, L. E. 2009.
658 Guidelines for the nomenclature of the human heat shock proteins. *Cell Stress and*
659 *Chaperones*, 14, 105-111.
660

661 KRAUSE, M., HECK, T. G., BITTENCOURT, A., SCOMAZZON, S. P.,
662 NEWSHOLME, P., CURI, R. & HOMEM DE BITTENCOURT, P. I. 2015. The
663 Chaperone Balance Hypothesis: The Importance of the Extracellular to Intracellular
664 HSP70 Ratio to Inflammation-Driven Type 2 Diabetes, the Effect of Exercise, and the
665 Implications for Clinical Management. *Mediators of inflammation*, 2015.
666

667 KREGEL, K. C. 2002. Invited review: heat shock proteins: modifying factors in
668 physiological stress responses and acquired thermotolerance. *Journal of applied*
669 *physiology*, 92, 2177-2186.

670

671 KUKREJA, R., KONTOS, M., LOESSER, K., BATRA, S., QIAN, Y.-Z., GBUR, C.,
672 NASEEM, S., JESSE, R. & HESS, M. 1994. Oxidant stress increases heat shock
673 protein 70 mRNA in isolated perfused rat heart. *American Journal of Physiology-*
674 *Heart and Circulatory Physiology*, 267, H2213-H2219.

675

676 LANCASTER, G., MØLLER, K., NIELSEN, B., SECHER, N. H., FEBBRAIO, M.
677 A. & NYBO, L. 2004. Exercise induces the release of heat shock protein 72 from the
678 human brain in vivo. *Cell stress & chaperones*, 9, 276.

679

680 LEE, B. J., EMERY-SINCLAIR, E. L., MACKENZIE, R. W., HUSSAIN, A.,
681 TAYLOR, L., JAMES, R. S. & THAKE, C. D. 2014. The impact of submaximal
682 exercise during heat and/or hypoxia on the cardiovascular and monocyte HSP72
683 responses to subsequent (post 24 h) exercise in hypoxia. *Extreme physiology &*
684 *medicine*, 3, 15.

685

686 LEE, B. J., MACKENZIE, R. W., COX, V., JAMES, R. S. & THAKE, C. D. 2014.
687 Human Monocyte Heat Shock Protein 72 Responses to Acute Hypoxic Exercise after
688 3 Days of Exercise Heat Acclimation. *BioMed Research International*.

689 LINDQUIST, S. & CRAIG, E. 1988. The heat-shock proteins. *Annual review of*
690 *genetics*, 22, 631-677.

691

692 LOCKE, M. & NOBLE, E. G. 1995. Stress proteins: the exercise response. *Canadian*
693 *journal of applied physiology*, 20, 155-167.

694

695 LUO, X., ZUO, X., ZHOU, Y., ZHANG, B., SHI, Y., LIU, M., WANG, K.,
696 MCMILLIAN, D. R. & XIAO, X. 2008. Extracellular heat shock protein 70 inhibits
697 tumour necrosis factor-alpha induced proinflammatory mediator production in
698 fibroblast-like synoviocytes. *Arthritis Res Ther*, 10, R41.

699

MARFELL-JONES, M., OLDS, T., STEWART, A. & CARTER, J. 2006. ISAK: Potchefstroom: International Standards for Anthropometric Assessment.

MORAN, D. S., SHITZER, A. & PANDOLF, K. B. 1998. A physiological strain index to evaluate heat stress. *American Journal of Physiology-Regulatory, Integrative and Comparative Physiology*, 275, R129-R134.

MORTEZA, A., NAKHJAVANI, M., LARRY, M., NARGESI, A. A. & ESTEGHAMATI, A. 2013. Heat shock protein 70 and albuminuria in patients with type 2 diabetes: a matched case control study. *Cell Stress and Chaperones*, 18, 815-819.

MORTON, J. P., MACLAREN, D. P., CABLE, N. T., BONGERS, T., GRIFFITHS, R. D., CAMPBELL, I. T., EVANS, L., KAYANI, A., MCARDLE, A. & DRUST, B. 2006. Time course and differential responses of the major heat shock protein families in human skeletal muscle following acute nondamaging treadmill exercise. *Journal of Applied Physiology*, 101, 176-182.

MOSELEY, P. L. 1997. Heat shock proteins and heat adaptation of the whole organism. *Journal of Applied Physiology*, 83, 1413-1417.

NAJAFIZADEH, S. R., GHAZIZADEH, Z., NARGESI, A. A., MAHDAVI, M., ABTAHI, S., MIRMIRANPOUR, H. & NAKHJAVANI, M. 2015. Analysis of serum heat shock protein 70 (HSPA1A) concentrations for diagnosis and disease activity monitoring in patients with rheumatoid arthritis. *Cell Stress and Chaperones*, 1-7.

OGAWA, K., KIM, H.-K., SHIMIZU, T., ABE, S., SHIGA, Y. & CALDERWOOD, S. K. 2012. Plasma heat shock protein 72 as a biomarker of sarcopenia in elderly people. *Cell Stress and Chaperones*, 17, 349-359.

PATEL, B., KHALIQ, A., JARVIS-EVANS, J., BOULTON, M., ARROL, S., MACKNESS, M. & MCLEOD, D. 1995. Hypoxia induces HSP 70 gene expression in human hepatoma (HEP G2) cells. *Biochemistry and molecular biology international*, 36, 907-912.

734 PERIARD, J., PATRICIA, R., CAILLAUD, C. & THOMPSON, M. Plasma Hsp72
735 and Hsp27 during moderate and intense exercise to exhaustion in the heat.
736 Proceedings of The Physiological Society, 2012. The Physiological Society.
737

738 PÉRIARD, J. D., RUELL, P., CAILLAUD, C. & THOMPSON, M. W. 2012. Plasma
739 Hsp72 (HSPA1A) and Hsp27 (HSPB1) expression under heat stress: influence of
740 exercise intensity. *Cell Stress and Chaperones*, 17, 375-383.
741

742 POCKLEY, A., SHEPHERD, J. & CORTON, J. 1998. Detection of heat shock
743 protein 70 (Hsp70) and anti-Hsp70 antibodies in the serum of normal individuals.
744 *Immunological investigations*, 27, 367-377.
745

746 POCKLEY, A. G., GEORGIADES, A., THULIN, T., DE FAIRE, U. &
747 FROSTEGÅRD, J. 2003. Serum heat shock protein 70 levels predict the development
748 of atherosclerosis in subjects with established hypertension. *Hypertension*, 42, 235-
749 238.
750

751 RITOSSA, F. 1962. A new puffing pattern induced by temperature shock and DNP in
752 *Drosophila*. *Experientia*, 18, 571-573.

753 RODRIGUES-KRAUSE, J., KRAUSE, M., O'HAGAN, C., DE VITO, G.,
754 BOREHAM, C., MURPHY, C., NEWSHOLME, P. & COLLERAN, G. 2012.
755 Divergence of intracellular and extracellular HSP72 in type 2 diabetes: does fat
756 matter? *Cell Stress and Chaperones*, 17, 293-302.
757

758 RUELL, P. A., SIMAR, D., PÉRIARD, J. D., BEST, S., CAILLAUD, C. &
759 THOMPSON, M. W. 2014. Plasma and lymphocyte Hsp72 responses to exercise in
760 athletes with prior exertional heat illness. *Amino acids*, 46, 1491-1499.
761

762 SANDSTRÖM, M. E., SIEGLER, J. C., LOVELL, R. J., MADDEN, L. A. &
763 MCNAUGHTON, L. 2008. The effect of 15 consecutive days of heat-exercise
764 acclimation on heat shock protein 70. *Cell Stress and Chaperones*, 13, 169-175.
765

766 SCHICK, C., ARBOGAST, M., LOWKA, K., RZEPKA, R. & MELCHERS, I. 2004.
 767 Continuous enhanced expression of Hsc70 but not Hsp70 in rheumatoid arthritis
 768 synovial tissue. *Arthritis & Rheumatism*, 50, 88-93.
 769
 770 SELKIRK, G. A., MCLELLAN, T. M., WRIGHT, H. E. & RHIND, S. G. 2009.
 771 Expression of intracellular cytokines, HSP72, and apoptosis in monocyte subsets
 772 during exertional heat stress in trained and untrained individuals. *American Journal of*
 773 *Physiology-Regulatory, Integrative and Comparative Physiology*, 296, R575-R586.
 774
 775 TAYLOR, L., MIDGLEY, A. W., CHRISMAS, B., MADDEN, L. A., VINCE, R. V.
 776 & MCNAUGHTON, L. R. 2010. Daily quadratic trend in basal monocyte expressed
 777 HSP72 in healthy human subjects. *Amino acids*, 38, 1483-1488.
 778
 779 WALSH, R., KOUKOULAS, I., GARNHAM, A., MOSELEY, P., HARGREAVES,
 780 M. & FEBBRAIO, M. 2001. Exercise increases serum Hsp72 in humans. *Cell stress*
 781 *& chaperones*, 6, 386.
 782
 783 WEINSTEIN, P. R., HONG, S. & SHARP, F. R. 2004. Molecular identification of the
 784 ischemic penumbra. *Stroke*, 35, 2666-2670.
 785 WHITHAM, M. & FORTES, M. 2006. Effect of blood handling on extracellular
 786 Hsp72 concentration after high-intensity exercise in humans. *Cell stress &*
 787 *chaperones*, 11, 304.
 788
 789 WHITHAM, M. & FORTES, M. B. 2008. Heat shock protein 72: release and
 790 biological significance during exercise. *Front Biosci*, 13, 1328-1339.
 791
 792

Table 1. Mean and peak physiological and thermoregulatory responses to each trial.

Data are mean $\pm$ SD.

Variable	Experimental group			ANOVA
	HOT70	HOT50	HYP50	Trial Main effect
Exercise duration (mins)	54.0 $\pm$ 9.4 ^{***}	60.0 $\pm$ 0.0	60.0 $\pm$ 0.0	F = 4.25, p = 0.032
Mean HR (beats.min ⁻¹)	180 $\pm$ 8 ^{***}	153 $\pm$ 12	149 $\pm$ 13	F = 14.17, p < 0.0001
Peak HR (beats.min ⁻¹)	190 $\pm$ 7 ^{***}	165 $\pm$ 10	157 $\pm$ 15	F = 14.61, p < 0.0001
Mean T _{rectal} (°C)	38.4 $\pm$ 0.3 [#]	38.0 $\pm$ 0.2	37.9 $\pm$ 0.4	F = 4.41, p = 0.028
Peak T _{rectal} (°C)	39.3 $\pm$ 0.4 [#]	38.7 $\pm$ 0.3 [#]	38.2 $\pm$ 0.4	F = 15.62, p < 0.0001
Mean PSI (AU)	7.2 $\pm$ 0.6 [*]	5.2 $\pm$ 1.0	4.8 $\pm$ 1.1	F = 11.81, p = 0.001
Peak PSI (AU)	9.2 $\pm$ 0.8 [*]	7.4 $\pm$ 1.2	5.9 $\pm$ 1.4	F = 13.74, p < 0.0001
Delta change in T _{rectal}	+2.5 $\pm$ 0.3 ^{***}	+1.4 $\pm$ 0.4 [#]	+0.8 $\pm$ 0.4	F = 34.34, p < 0.0001
Rate of T _{rectal} change (°C h ⁻¹)	2.9 $\pm$ 0.7 ^{***}	1.4 $\pm$ 0.4 [#]	0.8 $\pm$ 0.4	F = 31.79, p < 0.0001
AUC for T _{rectal} 38.5°C (°C min ⁻¹)	7.7 $\pm$ 4.1 ^{***}	1.3 $\pm$ 2.9	1.0 $\pm$ 2.7	F = 8.95, p = 0.002
AUC for T _{rectal} 39.0°C (°C min ⁻¹)	1.2 $\pm$ 0.9 [#]	0.0 $\pm$ 0.0	0.0 $\pm$ 0.0	F = 11.46, p = 0.001
Duration above 38.5°C (mins)	17.3 $\pm$ 9.7 [#]	5.3 $\pm$ 7.5	3.1 $\pm$ 8.8	F = 5.06, p = 0.018
Duration above 39.0°C (mins)	7.4 $\pm$ 3.9 ^{***}	1.4 $\pm$ 3.8	0.0 $\pm$ 0.0	F = 11.29, p = 0.001

Total n = 21, HOT70 n = 6, HOT50 n = 7; HYP50 n = 8.

AUC = Area under the curve. PSI = Physiological strain index.

*** denotes significantly different from HOT50 and HYP50 (p < 0.001)

denotes significantly different from HYP50 group (p < 0.05)

* denotes significantly different from HOT50 (p < 0.05)

Figure Legends

Figure 1. Standard curves (mean $\pm$ SD; $n = 4$) generated in duplicate from the same pre-prepared standards analyzed on the same 96 well plate. The amplification steps produce a clear increase in the assays sensitivity when compared to the EKS-715 kit reagents allowing for determination of low levels of eHSP70 in plasma samples.

Figure 2. Panel A and B display the optical density and eHSP70 values obtained from each assay. EKS-715 was able to detect eHSP70 in 6 of the 32 resting observations ($3.57 \pm 2.68 \text{ ng}\cdot\text{mL}^{-1}$), whereas ENZ-KIT measured eHSP70 in all 32 resting observations ($1.54 \pm 3.19 \text{ ng}\cdot\text{mL}^{-1}$). When data for an individual was available from both assays ($n = 12$, Panel C and D) the ENZ-KIT tended to indicate lower values, though this was not statistically significant ($p = 0.501$; $R^2 = 0.73$). Between test reliability for samples assayed on two different occasions was high (Panel E), with the between test CV 7.86% and a correlation coefficient of 0.99 (Panel F).

Figure 3. eHSP70 was detected at all-time points in each trial using the ENZ-KIT, and was only elevated post exercise in HOT70 (Panel A, $n = 6$), with no post exercise change in eHSP70 expression observed after HOT50 (Panel B, $n = 7$) and HYP50 (Panel C, $n = 8$). Individual data are shown with bars representing mean eHSP70.