Citation for published version:

Mellon, S, Grammatopoulos, G, Pandit, H, Murray, DW & Gill, HS 2014, 'Asymmetrical hip loading correlates with metal ion levels in patients with metal-on-metal hip resurfacing during sit-to-stand', *Hip International*, vol. 24, no. 1, pp. 20-26. <https://doi.org/10.5301/hipint.5000115>

DOI:

[10.5301/hipint.5000115](https://doi.org/10.5301/hipint.5000115)

Publication date:

2014

Document Version

Peer reviewed version

[Link to publication](#)

University of Bath

Alternative formats

If you require this document in an alternative format, please contact:
openaccess@bath.ac.uk

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

TITLE: Asymmetrical hip loading correlates with metal ion levels in patients with metal-on-metal hip resurfacing during sit-to-stand

Stephen J. Mellon¹, George Grammatopoulos¹, Hemant Pandit¹, David W. Murray¹, Harinderjit S. Gill²

1. University of Oxford, Nuffield Department of Orthopaedics, Rheumatology & Musculoskeletal Sciences.
2. Department of Mechanical Engineering, University of Bath, United Kingdom

Corresponding Author: Dr. Stephen Mellon

Email: stephen.mellon@ndorms.ox.ac.uk

Address: Botnar Research Centre, Nuffield Orthopaedic Centre, Oxford OX3 7LD

Tel: (+44) 01865 227454

ABSTRACT

The occurrence of pseudotumours following metal-on-metal hip resurfacing arthroplasty (MoMHRA) has been associated with high serum metal ion levels and consequently higher than normal bearing wear. Measuring ground reaction force is a simple method of collecting information on joint loading during a sit-to-stand (STS). We investigated vertical ground reaction force (VGRF) asymmetry during sit-to-stand for 12 MoMHRA patients with known serum metal ion levels. Asymmetry was assessed using two methods: a ratio of VGRF for implanted/unimplanted side and an absolute symmetry index (ASI). It was found that subjects with high serum metal ion levels preferentially loaded their implanted sides. The difference between the two groups was most apparent during the first 22% of STS. VGRF ratio showed significant and strong correlation with serum metal ion levels (Spearman's $\rho=0.8$, $p=0.003$). These results suggest that individual activity patterns play a role in the wear of MoMHRA and preferential loading of an implanted limb during the initiation of motion may increase the wear of metal-on-metal hip replacements.

INTRODUCTION

The soft tissue reactions observed in some patients with MoMHRA have been referred to by a number of terms such as adverse reaction to metal debris (ARMD¹), aseptic lymphocytic vasculitis associated lesions (ALVAL²), adverse local tissue reaction (ALTR³) and pseudotumour⁴. The term pseudotumour will be used in this study. Pseudotumours can be extremely destructive, causing one or a number of symptoms such as pain, spontaneous dislocation, pathological fracture and nerve palsy⁴.

The currently available data would suggest that the occurrence of pseudotumours is related to higher than normal wear, and pseudotumour is probably a response to the particulate wear debris⁵. Almost all MoMHRA implants are manufactured from cobalt-chrome alloy and levels of serum Co and Cr ions have been shown to be indicative of the amount of wear of MoMHRA implants⁶⁻⁸.

Information on joint loading can be inferred from ground reaction force (GRF). Studies measuring telemetered hip reaction forces from instrumented prosthesis have demonstrated, in terms of timing, that peak joint reaction force are concomitant with peak GRF⁹. A number of studies have therefore used measures determined from the GRF to assess outcomes following arthroplasty, for subjects performing gait¹⁰⁻¹⁵ and sit-to-stand^{11, 15, 16} activities. The aim of the current study was to examine the relationship between serum metal ion levels and the vertical component of the GRF (VGRF) symmetry during sit to stand activity for patients who had undergone MoMHRA.

METHODS: Subjects

One hundred and fifty eight (201 hips) MoMHRA patients had their serum metal ion levels measured in an on-going study, which was approved by the local ethics committee. Blood samples were collected in accordance with a previously described protocol¹⁷. Serum levels of cobalt and chromium were determined using inductively-coupled plasma mass spectrometry (ELAN DRC II, PerkinElmer Life and Analytical Sciences, Shelton, CT, USA) at the Laboratory of Clinical Biology, University Hospital Ghent, Belgium. Subjects were selected from this cohort for the current study according to their measured ion levels. All patients with high ion levels were invited to participate and those that accepted our invitation were recruited to the study. Thereafter, a number case controls (low-ion patients) matched for gender, age, BMI, prosthesis manufacturer and size were invited to participate. Those that accepted the invitation were recruited to the study. Twelve subjects with unilateral MoMHRA and a healthy unimplanted hip participated in the current study (7 females and 5 males, with a mean age of 59 (± 11) years a mean BMI of 24.5 (± 2.9) and a mean follow-up since surgery of 4.3 (± 1.1) years. These subjects had either a unilateral Birmingham Hip Resurfacing (BHR) (Smith and Nephew, Birmingham, UK) ($n=9$) or a unilateral Conserve Plus (Wright Medical Technology, Memphis, TN, USA) ($n=3$). The subjects were divided into two groups (Low Ions, $n=6$, and High Ions, $n=6$) based on their serum metal ion levels (Table 1). Serum metal ion levels were considered high if they exceeded 4.4 $\mu\text{g/l}$ for chromium or 4 $\mu\text{g/l}$ for cobalt¹⁸. There was no statistically significant difference between the groups in terms of height ($p > 0.05$) despite a 10 cm difference in the means.

METHODS: Data Acquisition and Processing

Motion analysis studies were performed in a dedicated motion analysis laboratory. The subjects were seated on a bench of height 550 mm with each foot on a force platform. Subjects were then asked to stand with their arms folded across their chests in order to obtain repeatable motions between subjects. Ground reaction forces were recorded at 1000 Hz using two force platforms; an AMTI OR6-1-1000 and an AMTI OR6-6-1000 (AMTI, MA, USA) were used. The output of each force platform was passed through a dedicated amplifier (MSA-6, AMTI, MA, USA) and data were collected using a VICON MX system (Vicon Ltd, Oxford, UK). The sit-to-stand activity was performed at least three times by each subject.

Post-processing was performed in Matlab (R2010, The MathWorks Inc., Natick, MA, USA). In the current study, the activity duration was defined from the total VGRF ($=VGRF_{\text{implanted}} + VGRF_{\text{unimplanted}}$). The activity start point was determined manually as automated methods for defining this point were found to be unreliable. A point on the total VGRF versus time characteristic of what is universally described as the loading phase was chosen (point A in Figure 1). The end point was defined automatically when the total VGRF crossed a threshold level, corresponding to the subject weight, for a second time. This period was chosen as it coincides with maximum force production^{11, 19, 20} and maximum hip contact force⁹. The total VGRF was normalised by body weight (% BW) and the duration was converted to a percentage of the STS activity to enable comparative analysis between subjects.

METHOD: VGRF Symmetry

VGRF symmetry was assessed using two methods: By deriving a VGRF ratio and using an absolute symmetry index^{21, 22}. The VGRF ratio was calculated by dividing

the VGRF for the implanted limb by that for the unimplanted limb (Equation 1). A VGRF ratio value of 1 indicated perfect symmetry, and values greater than 1 indicated a greater proportion of weight-bearing on the implanted limb.

$$VGRF\ Ratio = \frac{VGRF_{implanted}}{VGRF_{unimplanted}} \quad (1)$$

A symmetry index, originally described by Herzog *et al.* (1989) and subsequently modified to an absolute symmetry index (ASI)²¹⁻²³, was also calculated (Equation 2).

$$ASI = \left| \frac{2(VGRF_{implanted} - VGRF_{unimplanted})}{(VGRF_{implanted} + VGRF_{unimplanted})} \right| \times 100\% \quad (2)$$

An ASI of 0% represents perfect symmetry whereas the suggested level of acceptance for gait and stair activity are values less than 5% and 15%, respectively²⁴. As no threshold has been determined for STS, an ASI of 10% was adopted to determine acceptable levels of symmetry. The point of maximum VGRF and the associated VGRF ratio and ASI were recorded for each subject. The difference in VGRF ratio and ASI for Low versus High groups over the STS cycle was examined using the Mann-Whitney U test. The subjects were also divided by gender and the difference in VGRF ratio and ASI over the STS cycle was examined using the Mann-Whitney U test.

For the periods where significant differences were found between implanted and unimplanted limbs, the correlations between metal (chromium and cobalt) ion levels and the symmetry indicators were investigated using Spearman's rho. Statistical analyses were performed using PASW Statistics 18 (IBM, Armonk, NY, USA).

RESULTS: VGRF Symmetry

The values of VGRF ratio for the Low group appeared to remain symmetrical throughout STS, whereas the High group showed a tendency to preferentially load the implanted limb from 7% to 20% of STS (Figure 2A). The difference between the Low and the High group during this period was found to be statistically significant ($p < 0.05$, Figure 3A). The mean VGRF for the High group during the period between 0 and 22% STS was 15.1% BW and the mean VGRF ratio was 1.44. Between 23% and 35% STS the VGRF ratio for the High group switched from loading the implanted side to the unimplanted limb. However, the difference in VGRF ratio between the Low and the High groups for the period 35% to 100% did not show any statistical significance ($p > 0.1$). The difference in VGRF ratio between male and female patients showed no statistical significance ($p > 0.1$) at any point over the STS cycle.

The mean ASI for the Low and the High groups showed slightly different results to the VGRF ratio. The ASI of the Low group suggested that the VGRF was symmetrical (less than 10% threshold) between 6% to 69% STS. The VGRF of the High group showed greater asymmetry than the Low group over the whole of STS. There was a statistically significant ($p < 0.05$) difference between the two groups from 9% to 17% STS (Figure 3B). The ASI of the High group was under the 10% threshold in the period between 19% and 32% STS. This asymmetry was during the period when loading was transferred from the implanted to the unimplanted side. The results indicate that the subjects in the High group were preferentially loading their implanted side whereas the lower limb loading of the Low group was symmetrical. The ASI confirmed the asymmetry in VGRF for the High group but also suggested that the Low group were on the verge of asymmetry (9.9% ASI). The difference in ASI

between male and female patients showed no statistical significance ($p>0.1$) at any point over the STS cycle.

There were significant ($p<0.001$) and strong correlations between VGRF ratio and the levels of metal ions (Table 2); the correlations between ASI and metal ion levels were not significant. One subject in the High group was an extreme outlier, with a cobalt level of 87.6 $\mu\text{g/l}$ and a chromium level of 113.0 $\mu\text{g/l}$; the data for this subject was excluded to avoid skewing the correlation analysis. The correlations were strongest between VGRF ratio between 0 to 22% STS cycle (Figure 4); for chromium Spearman's rho was 0.802 ($p=0.003$) and for cobalt rho was 0.836 ($p=0.001$).

DISCUSSION

Subjects who have undergone MoMHRA show asymmetry in their VGRF during STS, and the degree of asymmetry correlates to the systemic metal ion levels measured in these individuals. Subjects with high systemic metal ion levels preferentially loaded their implanted limbs during the initial part of the STS. Metal ion levels are now accepted as surrogate measures of *in vivo* wear of metal-on-metal hip replacements. Whilst there are undoubtedly a considerable number of factors which influence the amount of wear of MoM implants, the implication of this study's findings is that loading during sit to stand activity has an influence on the function of the bearing. The extent of asymmetry in subjects who had low serum metal ion levels and high serum ion levels was different in terms of magnitude and the side that was loaded. The difference between the High and Low groups was most apparent during the first 22% of STS. During this period the High group preferentially loaded their implanted side, whereas the Low group loaded their limbs symmetrically.

It has been shown that THR patients spend 43% of time sitting, whereas standing and gait only take up 24.5% and 10% of time respectively²⁵. In addition it was found that sit-to-stand is performed approximately 50 times per day in patients with total joint replacement²⁶. It has been shown with hip simulators that under start-up and stopping conditions, the wear of MoM implants is increased²⁷. In the current study it has been shown that serum metal ion levels significantly and strongly correlate with VGRF asymmetry during the first 20% of STS for subjects with MoMHRA. This, therefore, suggests that joint loading asymmetry during the initiation of a frequently performed activity such as STS can lead to an increase in wear. Although the magnitude of asymmetry was relatively small the repetitious nature of the activity means there is an increased potential for increased bearing wear. The exact causes of asymmetry during STS are not clear. However, patients with intrinsic motion patterns that contribute to asymmetrical joint loading could be potentially educated to alter these patterns during rehabilitation.

After a prolonged period of rest, such as sitting for a length of time, there is probably a loss of fluid from between the bearing surfaces of a MoM implant. Relative motion between the bearing surfaces is needed to entrain fluid into the joint; it is not clear what period of time is required for establishment of a fluid film. It is now appreciated that the development of a fluid film is dependent upon the entrainment wedge, and this explains the findings of increased wear for steeply implanted acetabular components⁸. We speculate that for optimum bearing performance a balance is needed between the development of sufficient relative speed of the bearing surfaces to develop adequate lubrication and the application of load across the bearing

surfaces. We suggest that individuals who preferentially load their implanted limbs during the initiation of sit-to-stand will begin applying load before substantial lubrication is established, and this may explain the correlations found by this study.

Metal-on-metal implants probably represent the class of devices that have undergone the most extensive pre-clinical testing. The complications that have emerged with these devices were not predicted by extensive pre-clinical testing. One conclusion that can be drawn from recent experiences is that understanding of hip joint loading and wear is not as complete as was thought when these devices were designed. The pseudotumour issue should lead the orthopaedic community to re-examine the theoretical and experimental pillars upon which our understanding is built. The results from the current study suggest that activities which have not previously been considered important in pre-clinical testing may play a significant role in determining the wear performance of hip replacement bearings. Furthermore, variability in how an activity is performed can influence the amount of wear experienced.

The current study has a number of limitations. Rising from a chair in a “hands-free” manner may not reflect the real activity. Studies that have previously examined GRF asymmetry during STS have incorporated methods for eliminating the effects of the subjects’ dominant side²⁸, preferred foot position and seat height²⁹. This was not carried out in the current study and represents a limitation. In this study only the STS activity was considered and incorporating gait or stair-climbing may have given a clearer picture of the contribution of asymmetry to wear. The small number of subjects in this study also represents a limitation. There is a strong association between implant positioning and metal ion levels. In this study it was not possible to decouple the

relationship between implant positioning, metal ion levels and asymmetry in hip loading.

Subjects with MoMHRA display asymmetry in vertical ground reaction force during sit-to-stand. Asymmetry of VGRF at the beginning of STS correlates with metal ion levels measured in MoMHRA patients. This suggests that preferential loading of an implanted limb during the initial phase of sit-to-stand activity contributes to the wear of MoMHRA.

REFERENCES

1. D. J. Langton, S. S. Jameson, T. J. Joyce, N. J. Hallab, S. Natu, and A. V. F. Nargol: 'Early failure of metal-on-metal bearings in hip resurfacing and large-diameter total hip replacement', *Journal of Bone & Joint Surgery, British Volume*, 2010, **92-B**(1), 38-46.
2. H.-G. Willert, G. H. Buchhorn, I. Dipl, A. Fayyazi, R. Flury, M. Windler, G. Köster, and C. H. Lohmann: 'Metal-on-Metal Bearings and Hypersensitivity in Patients with Artificial Hip JointsA Clinical and Histomorphological Study', *The Journal of Bone & Joint Surgery*, 2005, **87**(1), 28-36.
3. T. P. Schmalzried: 'Metal-metal bearing surfaces in hip arthroplasty', *Orthopedics*, 2009, **32**(9).
4. H. Pandit, S. Glyn-Jones, P. McLardy-Smith, R. Gundle, D. Whitwell, C. L. M. Gibbons, S. Ostlere, N. Athanasou, H. S. Gill, and D. W. Murray: 'Pseudotumours associated with metal-on-metal hip resurfacings', *J Bone Joint Surg Br*, 2008, **90-B**(7), 847-851.
5. D. W. Murray, G. Grammatopoulos, R. Gundle, C. L. M. H. Gibbons, D. Whitwell, A. Taylor, S. Glyn-Jones, H. G. Pandit, S. Ostlere, H. S. Gill, N. Athanasou, and P. McLardy-Smith: 'Hip resurfacing and pseudotumour', *Hip International*, 2011, **21**(3), 279-283.
6. P. Campbell, P. E. Beaulé, E. Ebramzadeh, M. Leduff, K. D. Smet, Z. Lu, and H. C. Amstutz: 'The john charnley award: A study of implant failure in metal-on-metal surface arthroplasties', *Clinical Orthopaedics and Related Research*, 2006(453), 35-46.
7. K. De Smet, R. De Haan, A. Calistri, P. A. Campbell, E. Ebramzadeh, C. Pattyn, and H. S. Gill: 'Metal ion measurement as a diagnostic tool to identify problems with metal-on-metal hip resurfacing', *Journal of Bone and Joint Surgery - Series A*, 2008, **90**(SUPPL. 4), 202-208.
8. R. De Haan, C. Pattyn, H. S. Gill, D. W. Murray, P. A. Campbell, and K. De Smet: 'Correlation between inclination of the acetabular component and metal ion levels in metal-on-metal hip resurfacing replacement', *Journal of Bone and Joint Surgery - Series B*, 2008, **90**(10), 1291-1297.
9. G. Bergmann, G. Deuretzbacher, M. Heller, F. Graichen, A. Rohlmann, J. Strauss, and G. N. Duda: 'Hip contact forces and gait patterns from routine activities', *Journal of Biomechanics*, 2001, **34**(7), 859-871.

10. V. Lugade, A. Wu, B. Jewett, D. Collis, and L. S. Chou: 'Gait asymmetry following an anterior and anterolateral approach to total hip arthroplasty', *Clinical Biomechanics*, 2010, **25**(7), 675-680.
11. V. L. Talis, A. A. Grishin, I. A. Solopova, T. L. Oskanyan, V. E. Belenky, and Y. P. Ivanenko: 'Asymmetric leg loading during sit-to-stand, walking and quiet standing in patients after unilateral total hip replacement surgery', *Clinical Biomechanics*, 2008, **23**(4), 424-433.
12. R. N. Stauffer, G. L. Smidt, and J. B. Wadsworth: 'Clinical and Biomechanical Analysis of Gait Following Charnley Total Hip Replacement', *Clinical Orthopaedics and Related Research*, 1974, **99**, 70-77.
13. J. L. McCrory, S. C. White, and R. M. Lifeso: 'Vertical ground reaction forces: objective measures of gait following hip arthroplasty', *Gait & Posture*, 2001, **14**(2), 104-109.
14. R. M. Queen, T. S. Watters, A. N. Abbey, V. J. Sabesan, T. P. Vail, and M. P. Bolognesi: 'Gait Symmetry: A Comparison of Hip Resurfacing and Jumbo Head Total Hip Arthroplasty Patients', *The Journal of Arthroplasty*, 2011, **26**(5), 680-685.
15. R. L. Mizner and L. Snyder-Mackler: 'Altered loading during walking and sit-to-stand is affected by quadriceps weakness after total knee arthroplasty', *Journal of Orthopaedic Research*, 2005, **23**(5), 1083-1090.
16. M. C. Boonstra, B. W. Schreurs, and N. Verdonchot: 'The Sit-to-Stand Movement: Differences in Performance Between Patients After Primary Total Hip Arthroplasty and Revision Total Hip Arthroplasty With Acetabular Bone Impaction Grafting', *Physical Therapy*, 2011, **91**(4), 547-554.
17. S. J. MacDonald, W. Brodner, and J. J. Jacobs: 'A consensus paper on metal ions in metal-on-metal hip arthroplasties', *The Journal of Arthroplasty*, 2004, **19**(8, Supplement), 12-16.
18. C. D. S. Van Der Straeten, K; Grammatopoulos, G; Gill, H. S.: 'Interpretation of metal ions in unilateral and bilateral hip resurfacings', EORS, Vienna, Austria, 2011.
19. G. Roy, S. Nadeau, D. Gravel, F. Malouin, B. J. McFadyen, and F. Piote: 'The effect of foot position and chair height on the asymmetry of vertical forces during sit-to-stand and stand-to-sit tasks in individuals with hemiparesis', *Clinical Biomechanics*, 2006, **21**(6), 585-593.
20. F. Mourey, A. Grishin, P. D'Athis, T. Pozzo, and P. Stapley: 'Standing up from a chair as a dynamic equilibrium task: A comparison between young and elderly subjects', *Journals of Gerontology - Series A Biological Sciences and Medical Sciences*, 2000, **55**(9), B425-B431.
21. G. Giakas and V. Baltzopoulos: 'Time and frequency domain analysis of ground reaction forces during walking: an investigation of variability and symmetry', *Gait & Posture*, 1997, **5**(3), 189-197.
22. R. White, I. Agouris, R. D. Selbie, and M. Kirkpatrick: 'The variability of force platform data in normal and cerebral palsy gait', *Clinical Biomechanics*, 1999, **14**(3), 185-192.
23. W. Herzog, B. M. Nigg, L. J. Read, and E. Olsson: 'Asymmetries in ground reaction force patterns in normal human gait', *Med Sci Sports Exerc*, 1989, **21**(1), 110-114.
24. A. Stacoff, C. Diezi, G. Luder, E. Stussi, and I. A. Kramers-De Quervain: 'Ground reaction forces on stairs: effects of stair inclination and age', *Gait & Posture*, 2005, **21**(1), 24-38.
25. M. Morlock, E. Schneider, A. Bluhm, M. Vollmer, G. Bergmann, V. Müller, and M. Honl: 'Duration and frequency of every day activities in total hip patients', *Journal of Biomechanics*, 2001, **34**(7), 873-881.
26. I. Groot, H. Bussmann, H. Stam, and J. Verhaar: 'Small Increase of Actual Physical Activity 6 Months After Total Hip or Knee Arthroplasty', *Clinical Orthopaedics and Related Research*, 2008, **466**(9), 2201-2208.
27. G. E. Roter, J. B. Medley, J. D. Bobyn, J. J. Krygier, and F. W. Chan: 'Stop-dwell-start motion: a novel simulator protocol for the wear of metal-on-metal hip implants', in 'Tribology Series', (eds. M. P. G. D. D. Dowson, et al.), 367-376; 2002, Elsevier.

28. D. R. Burnett, N. H. Campbell-Kyureghyan, P. B. Cerrito, and P. M. Quesada: 'Symmetry of ground reaction forces and muscle activity in asymptomatic subjects during walking, sit-to-stand, and stand-to-sit tasks', *Journal of Electromyography and Kinesiology*, 2011, **21**(4), 610-615.
29. W. G. M. Janssen, H. B. J. Bussmann, and H. J. Stam: 'Determinants of the sit-to-stand movement: A review', *Physical Therapy*, 2002, **82**(9), 866-879.

Tables

Table 1. Subject details

GROUP	GENDER (M/F)	AGE	HEIGHT (m)	BMI	CHROME ($\mu\text{g/l}$)	COBALT ($\mu\text{g/l}$)	RANGE INCLINATION	RANGE ANTEVERSION
LOW	3/3	65 (± 4.9)	1.73 (± 0.08)	25 (± 2.9)	1.55 (± 1.1)	1.42 (± 0.6)	40.3° - 51.2°	13.7° - 30.34°
HIGH	4/2	53 (± 11.6)	1.66 (± 0.09)	24 (± 3.1)	24.3 (± 43.5)	19.73 (± 33.3)	34.6° - 62.0°	-9.9° - 34.1°

Table 2. VGRF and ASI correlation with metal ion levels: Spearman's rho (S.rho) and associated significance (Sig.)

Asymmetry Measure	Correlation Activity (%)	Chromium Levels		Cobalt Levels	
		S.rho	Sig.	S.rho	Sig.
VGRF ratio	0 – 22	0.77	p<0.001	0.82	p<0.001
ASI	9 – 17	0.47	p>0.1	0.46	p>0.1

FIGURE LEGENDS

Figure 1. Typical total vertical ground reaction force for sit to stand. Point A defines the start of the loading phase and was defined manually. Analysis was carried out from a start point 100 frames prior to point A. Point B was where the VGRF crossed a threshold defined by the subject's body weight (BW) for a second time and defines the analysis end point. The analysis duration (C) is shown by the shaded area.

Figure 2. (A) Mean VGRF ratio for the Low and the High groups. Ratios above 1 indicate greater loading on the implanted side. Error bars indicate standard deviation (B) Mean ASI for Low. Versus High. An ASI of 0 % represents perfect symmetry. Error bars indicate standard deviation

Figure 3. Box plot of VGRF ratio (A) and ASI (B) over the first 20% of STS for High versus Low groups.

Figure 4. Mean VGRF ratio for 0 -22% STS versus serum levels of chromium (Cr) and cobalt (Co) ($\mu\text{g/l}$)